



Basic UDI-DI & UDI-DI attributes

Basic UDI-DI set of data in UDI database

Principle: Each UDI-DI inherits the attributes of its linked Basic UDI-DI and devices DI

Basic UDI-DI

Applicable legislation (IVDR) (*)

- 2. Basic UDI-DI value (*)
- 2b Basic UDI-DI Issuing entity (*);
- 6. Manufacturer SRN (*)
- 5. Name and address of manufacturer
- 7. Name and address and SRN of AR
- 9. Risk class (*)
 - A.2.14 Intended for self-testing (Y/N) (*)
 - A.2.14 Intended for near-patient-testing (Y/N) (*)
- Companion diagnostics (Y/N) (*)
- Instrument(Y/N) (*)
- Reagent(Y/N) (*)
- Professional testing (Y/N) (*)
- 11. A. Name and/or, if applicable, device model that identifies the device(s) with this BASIC UDI-DI in the technical documentation and/or certificate or declaration of conformity
(Name and/or model shall be provided)

UDI-DIs

- 0. UDI-DI value (*)
- 0b. UDI-DI Issuing entity (*)
- Secondary DI (value and issuing entity)
- 11.B. Reference, Article or Catalogue number (*)
- Device with Direct marking (Y/N) (*)
- Direct marking UDI-DI value (*)
- Direct marking UDI-DI issuing entity (*)
- 1. Quantity of device(s) (*)
- 3. Type of UDI-PI (*)
- 4. Unit of use UDI-DI (*)
- 13. Storage/handling conditions
- 10-14. Name(s)/Trade name(s) (including languages)
- 12. Additional product description
- 19. URL for additional information
- 15. Labelled as single use (YN) (*)
- 16. Maximum number of reuse (*)
- 17. Device labelled sterile (Y/N) (*)
- 18. Need for sterilisation (Y/N) (*)
- 20. Critical warnings or contra-indications
- 8. Medical device nomenclature (CND) code (1)
- 21. Status
- 27 (A.2.10). In the case of devices designed and manufactured by another legal or natural person as referred in Article 10(14), the name, address and contact details of that natural/legal person

UDI-DIs (container package DI)

- 0. UDI-DI value (*)
- 0b. Issuing entity (*)
- 1. Quantity per package (*)
- 21. Status

(1) Nomenclature decision:

<https://ec.europa.eu/docroom/documents/34264>

(*) may not be changed

- Mandatory
- Mandatory if applicable
- Optional

Other Device Data attributes

Basic UDI-DI

- **A.2.2 Certificate IDs (with NB, type .. Link);**
- **A.2.11 SSP;**
- **A.2.9 Performance study IDs (..link);**
- **A.2.5 Presence of Human tissues/Cells (Y/N) (*);**
- **A.2.6 Presence of Animal tissues/Cells (Y/N) (*);**
- **A.2.7 Presence of Substances/cells of microbial origin (Y/N) (*);**
- **Kit (Y/N) (*);**




 Provided by NB or for certificate ID under Art 26(2) provided by manufacturer and confirmed by NB

Version April 2019

UDI-DIs

- **A.2.13 New Device (Y/N) (*);**
- **A.2.3 Member State of the Placing on the EU Market of the Device (*);**
- **A.2.4 Member State(s) where the Device is made available in the Country;**

(*) may not be changed

-  **Mandatory**
-  **Mandatory if applicable**
-  **Optional**